



Box R Water Analysis Laboratory

567 NW Second Street
Prineville, Oregon 97754
541-447-4911

Mr. Brad Sperry

June 24, 2016

c/o Fossil School District

PO Box 206

Fossil, OR 97830

Sample Nbr: X022603 - 610


Dear Mr. Sperry,

Attached is a copy of your drinking water – Lead – test results, sampled on June 9, 2016, at Fossil School Buildings, Fossil, OR. Your analysis was performed by Neilson Research Corp. in Medford, OR. Please do not hesitate to call Box R Water Analysis Laboratory with any questions you may have in regards to your water testing.

Thank you for using Box R Water Analysis Laboratory, we appreciate your business.

Sincerely,

Sherri K. Miyazaki – Box R Water Analysis Laboratory Director



NEILSON RESEARCH CORPORATION

Environmental Testing Laboratory

6/21/2016

Sherri Miyazaki
Box R Waterlab
567 NW Second Street
Prineville, OR 97754

TEL: (541) 447-4911
FAX (541) 447-4917

RE: X022603-610 Fossil School District

Order No.: 1606677

Dear Sherri Miyazaki:

Neilson Research Corporation received 8 sample(s) on 6/16/2016 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Alec C Smith
Project Manager

245 S Grape St • Medford, OR 97501 • (541) 770-5678
400 SE G St, Suite B • Grants Pass, OR 97526 • (541) 479-4053

www.nrclabs.com

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

CLIENT: Box R Waterlab
Project: X022603-610 Fossil School District
Lab Order: 1606677

Date: 21-Jun-16

CASE NARRATIVE

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: **1606677**
NRC Sample ID: **1606677-01A**
Collection Date: **6/9/2016 8:30:00 AM**
Received Date: **6/16/2016 10:55:00 AM**
Reported Date: **6/21/2016 11:44:44 AM**

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022603
Collectors Name: CGM
Sample Location: Grade School Kitchen Sink
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP		Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
		Accredited	Result						
Lead	EPA 200.8	A	0.0146		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: **1606677**
NRC Sample ID: **1606677-02A**
Collection Date: **6/9/2016 8:30:00 AM**
Received Date: **6/16/2016 10:55:00 AM**
Reported Date: **6/21/2016 11:44:45 AM**

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022604
Collectors Name: CGM
Sample Location: Kindergarten DF
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP		Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
		Accredited	Result						
Lead	EPA 200.8	A	0.00518		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606677
NRC Sample ID: 1606677-03A
Collection Date: 6/9/2016 8:30:00 AM
Received Date: 6/16/2016 10:55:00 AM
Reported Date: 6/21/2016 11:44:45 AM

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022605
Collectors Name: CGM
Sample Location: 1st-2nd Grade DF
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP		Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
		Accredited	Result						
Lead	EPA 200.8	A	0.0114		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606677
NRC Sample ID: 1606677-04A
Collection Date: 6/9/2016 8:30:00 AM
Received Date: 6/16/2016 10:55:00 AM
Reported Date: 6/21/2016 11:44:45 AM

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022606
Collectors Name: CGM
Sample Location: High School Hallway DF
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP		Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
		Accredited	Result						
Lead	EPA 200.8	A	0.00207		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: **1606677**
NRC Sample ID: **1606677-05A**
Collection Date: **6/9/2016 8:30:00 AM**
Received Date: **6/16/2016 10:55:00 AM**
Reported Date: **6/21/2016 11:44:45 AM**

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022607
Collectors Name: CGM
Sample Location: Lower Hallway DF
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP Accredited	Result	Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
Lead	EPA 200.8	A	0.00195		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606677
NRC Sample ID: 1606677-06A
Collection Date: 6/9/2016 8:30:00 AM
Received Date: 6/16/2016 10:55:00 AM
Reported Date: 6/21/2016 11:44:45 AM

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022608
Collectors Name: CGM
Sample Location: Home Ec Sink Near Door
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP Accredited	Result	Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
Lead	EPA 200.8	A	0.00188		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: 1606677
NRC Sample ID: 1606677-07A
Collection Date: 6/9/2016 8:30:00 AM
Received Date: 6/16/2016 10:55:00 AM
Reported Date: 6/21/2016 11:44:45 AM

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022609
Collectors Name: CGM
Sample Location: Home Ec Sink Near Window
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP Accredited	Result	Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
Lead	EPA 200.8	A	0.00389		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Box R Waterlab
567 NW Second Street
Prineville, OR 97754

Lab Order: **1606677**
NRC Sample ID: **1606677-08A**
Collection Date: **6/9/2016 8:30:00 AM**
Received Date: **6/16/2016 10:55:00 AM**
Reported Date: **6/21/2016 11:44:45 AM**

Sample Information:

X022603-610 Fossil School District

Client Sample ID: X022610
Collectors Name: CGM
Sample Location: High School Hallway DF
Source: City Water

ANALYTICAL RESULTS

Analyses	Method	NELAP		Qual	MRL	Units	EPA Limit	Date Analyzed	Analyst
		Accredited	Result						
Lead	EPA 200.8	A	0.00206		0.000103	mg/L	0.020 AL	6/17/2016	OML

Notes: ND - Not Detected at the MRL

N.L. - No Limit

MRL - Minimum Reporting Limit

Neilson Research Corporation

Date: 21-Jun-16

CLIENT: Box R Waterlab
Work Order: 1606677
Project: X022603-610 Fossil School District

ANALYTICAL QC SUMMARY REPORT

TestCode: ICPMS_200.8 SCHOOL

Sample ID	MB-35484	SampType: MBLK	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 6/17/2016	RunNo: 87908						
Client ID:	ZZZZZ	Batch ID: 35484	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 6/17/2016	SeqNo: 1311774						
Analyte		Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead		ND	0.000104	0.1	0	96.0	85	115				

Sample ID	LCS-35484	SampType: LCS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 6/17/2016	RunNo: 87908						
Client ID:	ZZZZZ	Batch ID: 35484	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 6/17/2016	SeqNo: 1311775						
Analyte		Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead		0.09600	0.000104	0.1	0	96.0	85	115				

Sample ID	1606679-02AMS	SampType: MS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 6/17/2016	RunNo: 87908						
Client ID:	ZZZZZ	Batch ID: 35484	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 6/17/2016	SeqNo: 1311792						
Analyte		Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead		0.09554	0.000104	0.1	0.0008454	94.7	70	130				

Sample ID	1606679-02AMSD	SampType: MSD	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 6/17/2016	RunNo: 87908						
Client ID:	ZZZZZ	Batch ID: 35484	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 6/17/2016	SeqNo: 1311793						
Analyte		Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead		0.09522	0.000104	0.1	0.0008454	94.4	70	130	0.09554	0.338	20	

Qualifiers: E Value above quantitation range H Holding times for preparation or analysis exceeded J Analyte detected below quantitation limits
 ND Not Detected at the Minimum Reporting Limit R RPD outside accepted recovery limits S Spike Recovery outside accepted recovery limits